

Microbial DNA Enrichment

MolYsis™ Basic10

Sample pre-treatment kit for background-free PCR analysis of whole blood and other liquid samples

Large Size Sample Volumes (5 to 10ml)

Kit includes all ingredients for the following steps of selective lysis of host cells and the degradation of released DNA:

- Lysis of human/animal cells
- Degradation of human/animal DNA
- Degradation of cell walls of Gram-positive and Gram-negative bacteria and fungi

To be used with other DNA isolation kits

- For research use only -



© 2014 Molzym, all rights reserved

Version 2.0

Date of first release: 11/2011

Last update: 10/2014

Contents

Kit Information	3
Kit Contents – <i>MolYsis™ Basic10</i>	3
Symbols	3
Storage and Stability	3
Product Use Limitations.....	3
Safety Information	4
Risk and Safety Phrases.....	4
Introduction	5
Kit Description	5
The <i>MolYsis™ Basic10</i> Technology	6
List of Strains detected	7
Recommendations for PCR Analysis of Bacteria	7
Protocol	8
How to Start.....	8
Procedure.....	9
A) Fill up procedure for samples 5ml to less than 10ml volume.....	9
B) Sample pre-treatment procedure.....	9
Supplementary Information	11
Troubleshooting.....	11
References.....	13
Information <i>MolYsis™ Complete</i> Kits and DNA-Free PCR Reagents.....	15
Technical Support.....	16
Order Information	16
Contact.....	16

Kit Information

Kit Contents – *MolYsis™ Basic10*

	50 reactions	100 reactions
Extraction Buffers (store at +18 to +25°C)		
<i>SU</i>	2x 125ml	4x 125ml
<i>CM</i>	2x 100ml	4x 100ml
<i>DB1</i>	2x 100ml	4x 100ml
<i>RS</i>	1x 50ml	2x 50ml
<i>RL</i>	1x 15ml	2x 15ml
Enzymes (store at -15 to -25°C)		
<i>MolDNase B</i> , solution	1x 0.5ml	2x 0.5ml
<i>BugLysis</i> , solution	1x 1.0ml	2x 1.0ml
<i>β-mercaptoethanol</i> , solution	1x 0.08ml	2x 0.08ml
Manual		
Manual	1x	1x

Symbols

Symbols used in labelling and in section 'Risk and Safety Phrases' (pages 4 to 5).



Content of the package

REF

Catalogue number



GHS05 – Corrosive



Manufactured by

Cont.

Consult instructions for use



GHS06 – Toxicity



Use by



Harmful, Xn / Irritant, Xi



GHS09 – Environmentally Hazardous



Temperature limitation (store at)



Toxic; T

LOT

Batch code



Hazardous to the environment; N

Storage and Stability

Guarantee for full performance of the kit as specified in this manual is only valid if storage conditions are followed. Please take care that *MolDNase B*, *BugLysis* and *β-mercaptoethanol* are handled and stored at -15 to -25°C. Store buffers at room temperature (+18 to +25°C).

Guarantee for full performance of the kit is given for up to **24 months** at the conditions specified.

Product Use Limitations

This product is for **research use only** and not for use in diagnostic procedures.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable sleeve covers, sterile disposable gloves and protective goggles. For more information, please consult the appropriate material safety data sheets (MSDS) which are available on request.

CAUTION: Never add hypochlorite (bleach) or acidic solutions directly to the sample-preparation waste.

Buffer *CM* contains guanidine hydrochloride, which can form highly reactive compounds and toxic gases when combined with hypochlorite or other acidic solutions. If liquid containing these buffers is spilt, clean with suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 70% (v/v) ethanol. This kit is to be used only by skilled personnel trained for handling infectious material and molecular biological methods. To avoid false analytic results by DNA contamination of reagents and infection of the user by infectious agents during handling, always wear sterile protective gloves, disposable sleeve covers, a lab coat and protective goggles. Work in a Class II biological safety cabinet irradiated with UV before starting according to the instruction manual of the manufacturer. The UV lamp must be switched off during working. Follow the instructions of the manufacturer for maintenance of the workstation. Dispose potentially infectious material and the waste of the sample preparation according to the national directive of the health organisation (e.g. Richtlinie über die ordnungsgemäße Entsorgung von Abfällen aus Einrichtungen des Gesundheitsdienstes, 2002).

Separate Material Safety Data Sheets for chemicals used are available on request. Molzym reserves the right to alter or modify the product to improve the performance of the kit.

Risk and Safety Phrases



Buffer *SU* and *RL*

Contains sodium azide (<1%): **Harmful**.
Risk and safety phrases^{*(page 5)}: **R22-32, S28-45-60-61**



Buffer *CM*

Contains guanidine hydrochloride (>25%): **Harmful, irritant**.
Risk and safety phrases^{*(page 5)}: **R22-36/38, S26**

2-mercaptoethanol (β -*mercaptoethanol*):

Poisonous, irritating, environmental hazardous



Directive 67/548/EWG and 1999/45/EG

Risk and safety phrases^{*(page 5)}: **R23/24/25-38-41-50/53, S26-36/37/39-45-61**



Regulation (EC) No. 1272/2008



Danger

Hazard and precautionary statements^{** (page 5)}: **H227-H301-H310+H330-H315-H318-H410; P273-P301+P310-P302+P352-P304+P340-P305+P351+P338**

Emergency Information (24-hours service)

Emergency medical information in English, French, and German can be obtained 24 hours a day from: Poison Information Centre Mainz, Germany; Tel: +49(0)6131 19240 Outside of Germany: Please contact the regional company representation in your country.

* **R22**: Harmful if swallowed; **R23/24/25**: Toxic by inhalation, in contact with skin and if swallowed; **R32**: Contact with acids liberates very toxic gas; **R36/38**: Irritating to eyes and skin; **R38**: Irritating to skin; **R41**: Risk of serious damage to eyes **R50/53**: Very toxic to aquatic organisms, may cause long-term adverse effects in the aquatic environment;

S26: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Refer to special instructions/safety data sheet, **S28**: After contact with skin, wash immediately with plenty of water, **S36/37/39**: Wear suitable protective clothing, gloves and eye/face protection; **S45**: In case of accident or if you feel unwell seek medical advice immediately (show the label where possible); **S60**: This material and its container must be disposed of as hazardous waste; **S61**: Avoid release to the environment. Refer to special instructions/safety data sheet.

** **H227**: Combustible liquids; **H301**: Toxic if swallowed; **H310+H330**: Fatal if swallowed or in contact with skin; **H315**: Causes skin irritation; **H318**: Causes serious eye damage; **H410**: Very toxic to aquatic life with long lasting effects;

P273: Avoid release to the environment; **P301+P310**: IF SWALLOWED: Immediately call a POISON CENTER or doctor/physician; **P302+P352**: IF ON SKIN: Wash with plenty of soap and water; **P304+P340**: IF INHALED: Remove to fresh air and keep at rest in a position comfortable for breathing; **P305+P351+P338**: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing

Introduction

Kit Description

Molecular analysis of pathogenic bacteria and fungi in liquid samples from clinical materials and animal systems (e.g., blood and other body liquids) can be severely disturbed by a high background of host DNA. *MolYsis™ Basic10* removes this background of host DNA and thereby increases the reliability of the molecular analysis of pathogens in clinical and other samples. The kit contains all ingredients for the selective lysis of host cells and the degradation of released nucleic acids (DNA and RNA) in samples.

Patented *MolYsis™ Basic10* is a sample pre-treatment tool for the removal of host as well as dead cell DNA. The kit can be used as a module in conjunction with any other nucleic acid extraction kit designed for handling in the mini bind-wash-elute format (e.g., mini spin columns, automated systems). Molzym also supplies kits, manual *MolYsis™ Complete5* (≤1 and 5ml; D-321-050) and *MolYsis™ Complete10* (5 to 10ml; D-325-050) and automated *SelectNA Blood Pathogen Kit* (1ml; D-340-048), for the complete process of microbial DNA isolation from clinical and other samples, including sample pre-treatment, enrichment and lysis of bacterial and fungal cells, DNA extraction and DNA purification. Whereas other systems result in a mixture of host and microbial DNA, sample pre-treatment tool, *MolYsis™ Basic10*, enables the preparation of microbial DNA from samples. Only two steps are needed to obtain a sample that is depleted of host and dead cell DNA (Fig. 1, page 6):

- i) The addition of a chaotropic buffer to a sample lyses the host cells, whereas bacterial and fungal cells are unaffected.
- ii) The DNA released from host cells is degraded by Molzym's proprietary, chaotrope-resistant *MolDNase B*. Thereafter, pathogen cells are sedimented, treated with *BugLysis* reagent to degrade cell walls of Gram-negative and Gram-positive bacteria and fungi and then further processed by protocols for the extraction and purification of nucleic acids.

MolYsis™ Basic10 allows for the pre-treatment of 5 to 10ml liquid samples.

Samples evaluated:

Human origin: Whole blood (with anti-coagulants), synovial fluid, pleural fluid, cerebrospinal fluid, ascites fluid, pus, broncho-alveolar lavage, nasal douche fluid, urine

Animal origin: Whole blood (with anti-coagulants) from mouse, rat, and monkey, hamster ovary cell culture ($\leq 5 \times 10^8$ cells per sample), monkey renal cell culture.




Fig. 1: The principle of testing for bacterial and fungal pathogens in liquid samples by **MolYsis™ Basic10** and other Kits for the purification of pathogen DNA.

The **MolYsis™ Basic10** Technology

MolYsis™ Basic10 is Molzym's proprietary, patented technology enabling the pre-treatment of whole blood and other liquid clinical samples for depletion of human/animal DNA and enrichment of microbial organisms. The procedure includes protocols for human/animal DNA removal and universal lysis of Gram-negative and Gram-positive bacteria and fungi. A chaotropic buffer is added to a sample which lyses the human/animal cells (pathogens are unaffected) and the nucleic acids released are degraded by *MolDNase B*. Pathogen cells are subsequently centrifuged and treated by the *BugLysis* reagent for the degradation of cell walls. At the end, pathogen DNA is purified by other DNA isolation protocols including a proteinase digestion step. The preparation can then be used in broad-range and other PCR assays for analysis of pathogens.

List of Strains detected

The **MolYsis™ Basic10** technology has been evaluated with a variety of clinical samples (below). *BugLysis* reagent is a component of all kits and designed to lyse Gram-positive and Gram-negative bacteria and fungi with high efficiency. Strains from the following genera have been identified in clinical material so far (universal 16S PCR for bacteria, universal 18S PCR for fungi, plus sequencing), showing the broad range of lysing capability of *BugLysis*:

Gram-negative bacteria: *Acinetobacter*, *Aeromonas*, *Bacteroides*, *Bartonella*, *Bordetella*, *Borrelia*, *Bradyrhizobium*, *Brevibacterium*, *Candidatus Neoehrlichia*, *Citrobacter*, *Cloacibacterium*, *Coxiella*, *Dialister*, *Edwardsiella*, *Enterobacter*, *Escherichia*, *Fusobacterium*, *Haemophilus*, *Klebsiella*, *Leptotrichia*, *Methylobacterium*, *Moraxella*, *Morganella*, *Neisseria*, *Parabacteroides*, *Paracoccus*, *Petrobacter*, *Proteus*, *Providencia*, *Pseudomonas*, *Ralstonia*, *Raoultella*, *Serratia*, *Sphingomonas*, *Stenotrophomonas*, *Veillonella*, *Weeksella*, *Zoogloea*.

Gram-positive bacteria: *Actinomyces*, *Anaerococcus*, *Bacillus*, *Clostridium*, *Corynebacterium*, *Dolosigranulum*, *Enterococcus*, *Facklamia*, *Fingoldia*, *Gemella*, *Granulicatella*, *Lactobacillus*, *Lactococcus*, *Leifsonia*, *Listeria*, *Micrococcus*, *Mycobacterium*, *Nocardia*, *Parvimonas*, *Peptostreptococcus*, *Propionibacterium*, *Rothia*, *Ruminococcus*, *Staphylococcus*, *Streptococcus*, *Tropheryma*, *Vagococcus*.

Fungi: *Aspergillus*, *Candida*, *Cladosporium*, *Cryptococcus*, *Didymella*, *Davidiella*, *Malassezia*, *Peniophora*, *Saccharomyces*, *Shizophyllum*, *Sistotrema*, *Sporobolomyces*, *Udeniomyces*

Recommendations for PCR Analysis of Bacteria

Avoidance of DNA contamination: PCR analysis demands special care with respect to the avoidance of contamination from exogenous sources. Take care to separate places of DNA preparation from places where PCR reagents are handled, in particular preparation of mastermixes, pipetting into PCR tubes and performance of PCR runs. Wear sterile protective gloves at any handling step, also during DNA preparation. Frequently change sterile protective gloves during handling. Use only sterilized or, optimally, guaranteed DNA-free disposables. If analysis of bacteria is desired, e.g., identification by sequencing of broad-range 16S amplification products, it is important to make sure that only polymerases (e.g., *Taq* polymerase) free of DNA contamination are used. For this purpose, Molzym offers guaranteed DNA-free *MolTaq 16S* (P-019-0100). Also, Molzym offers a DNA-free mastermix (*Mastermix 16S Complete*; S-020-0100) containing primers for universal 16S rDNA amplification of bacterial sequences. Generally, for each analysis, run positive and negative controls to check for proper performance of the reaction and sterility of reagents and buffers used.

Call us for further information at +49(0)421 69 61 62 0.

Protocol

Large size sample pre-treatment for DNA isolation (5 to 10ml Liquid)

How to Start

Caution:

Work in a UV Class II biological safety cabinet. The UV lamp must be switched off during working. Use protective gloves and a disposable lab coat when handling infectious material!

! **Body liquid specimens:** Sampled under aseptic conditions and transferred to a sterile sample container (not supplied).

Whole blood samples: Use only EDTA or citrate-stabilized blood

! For optimal results, use only fresh samples. **Do not freeze samples** to avoid loss of pathogen DNA due to cell disruption.

! To be supplied by the user:

- 1x UV Class II biological safety cabinet
- 1x high speed centrifuge and fixed angle rotor for 50ml tubes (9,500xg)
- 1x bench top microcentrifuge ($\geq 12,000xg$)
- 1x thermomixer (2.0ml tubes)
- 1x vortexer
- 1x cooling rack for 1.5ml tubes (-15 to -25°C)
- Sample racks
- Precision pipettes and sterile filter pipette tips allowing pipetting volumes of up to 20 μ l, up to 200 μ l and up to 1000 μ l
- Sterile, disposable 5ml pipette equipped with aerosol filter, or a 5ml tip of a precision pipette.
- Sterile 50ml tubes (Cellstar tubes, order no. 227261, Greiner Bio-One / 50ml Centrifuge Tubes, Cat. no. 21008-242, VWR) for preparation of sample lysates by high speed centrifugation. If using other brands, **make sure that tubes can be used at RCF of 9,500xg**
- 2.0ml micro tubes, Biosphere®, Sarstedt, Germany (72.695.200) for bacterial and fungal cell lysis and DNA extraction
- 1x other Mini kit for nucleic acid extraction and purification

! Take care that *MolDNase B*, *BugLysis* and β -mercaptoethanol solutions are placed in a cooling rack adjusted to -15 to -25°C. Replace enzymes and reagent to the freezer (-15 to -25°C) immediately after handling.

! Adjust a thermomixer to 37°C

! To avoid contamination, close caps of bottles after removal of solution.

Approximate time for 4 parallel pre-treatments of samples:

80min (including 15min hands on time)

Procedure

A) Fill up procedure for samples 5ml to less than 10ml volume

Samples ≥ 5 and less than 10ml are filled up using buffer *SU*. Transfer the sample by pipetting into a sterile 50ml Falcon tube (not supplied, specification page 8). Then add buffer *SU* using a disposable 5ml pipette until reaching the 10ml mark of the tube. Discard the pipette/pipette tip with residual buffer *SU*. Continue with the pre-treatment procedure part B (below).

B) Sample pre-treatment procedure

1. **Pipette 10ml sample into a sterile 50ml Falcon tube (not supplied; specification, page 8) or use filled-up sample (part A, page 9). Then add 4ml buffer *CM*. Vortex at full speed for 15s. Let stand on the bench at room temperature (+18 to +25°C) for 5min.**

Buffer *CM* is a chaotropic buffer that lyses the human/animal cells. For optimal results it is important to mix thoroughly.

Caution: buffer *CM* is an irritant. Avoid contact with skin and eyes.

2. **Add 4ml buffer *DB1* and 10 μ l *MolDNase B* to the lysate and immediately vortex for 15s. Let stand on the bench for 15min.**

During this step the DNA released from human/animal cells are degraded.

3. **Centrifuge tube in a high speed centrifuge at 9,500xg for 10min. Thereafter, carefully decant the supernatant.**

4. **Add 1ml buffer *RS* and resuspend the sediment by vigorous vortexing.**

The pellet consists of cell debris and pathogen cells. Resuspension may take some time. Take care that all visible material has been resuspended.

5. **Transfer the suspension by pipetting to a sterile 2.0ml tube (not supplied; specification, page 8). Centrifuge tube in a bench top microcentrifuge ($\geq 12,000xg$) for 5min. Carefully remove the supernatant by pipetting and discard.**

This washing removes residual *MolDNase B* activity, chaotropic salts and most of the PCR inhibitors.

Note: At this point the procedure can be interrupted by freezing the sample (-15 to -25°C). For further processing, thaw the sample to room temperature (+18 to +25°C) and proceed with step 6 (page 10).

6. Add 80µl buffer *RL* and resuspend the sediment by vigorous vortexing.

Potential residual small particles in the suspension can be neglected, because they are dissolved during enzymatic treatments, in particular *Proteinase K* digestion with other kits (page 9).

7. Add 20µl *BugLysis* solution and 1.4µl *β-mercaptoethanol*, vortex for 15s and incubate tube in a thermomixer at 37°C and 1,000rpm for 30min.

The cell walls of potentially present bacteria and fungi are degraded.

Caution: *β-mercaptoethanol* is toxic. Take care not to inhale and otherwise come into contact with.

Further processing for DNA extraction and purification (other Mini kits):

After step 7, the microbial lysat can be processed by protocols for DNA isolation using manual or automated commercial kits or in-house extraction protocols. Protocols must include a protease or *Proteinase K* digestion after step 7 (above). Note that protease/ *Proteinase K* treatment is essential for optimal results. For this purpose, fill the microbial lysat (step 7, above) up to the sample volume of this kit with buffer *RL* (e.g., for 200µl sample volume add 100µl *RL* to the microbial lysat). For elution of the DNA from the column matrix, good experience was made using DNA-free water, PCR grade (Molzym order no. P-020-0003) heated to 70°C.

Supplementary Information

Troubleshooting

This guide may help solve problems that may arise. The Molzym team is always pleased to answer any of your questions about our products.

Phone: +49(0)421 69 61 62 0 • **E-Mail:** support@molzym.com

Observation	Possible cause	Comments/suggestions
Strong human/animal DNA background in gel electrophoresis or Real-Time PCR	<ul style="list-style-type: none"> • Buffer <i>CM</i> not added • Buffer <i>DB1</i> not added • <i>MolDNase B</i> not added • Solutions not mixed 	<p>Eluates usually contain traces of human/animal DNA co-eluted with bacterial/fungal DNA. If the extraction has not been performed according to the protocol, increased amounts of human/animal DNA can be the result, which negatively influences the PCR reaction. Ensure that buffer <i>CM</i> has been added to lyse human/animal cells. Accordingly, addition of buffer <i>DB1</i> and <i>MolDNase B</i> is obligate. Keep the <i>MolDNase B</i> vial chilled, because warming may reduce enzyme activity and hence increase human/animal DNA background. It is important that solutions are thoroughly mixed after addition of buffers. Follow instructions for vortexing.</p>
No pathogen DNA detectable (spiking test with negative blood)	<ul style="list-style-type: none"> • Insufficient lysis • Insufficient homogenisation • Pathogen titre too low • Loss of nucleic acids during purification 	<p>Make sure that <i>BugLysis</i> and <i>β-mercaptoethanol</i> treatments have been performed. Be aware that DNA is visible in a gelelectrophoresis only at amounts approx. >10ng (approx. >2x 10⁷ <i>E. coli</i> cells). Use PCR based procedures for detection and quantitation of bacteria <10⁷ cells.</p> <p>If the pellets from steps 4 and 6 (page 9) are not totally homogenized, bacterial and fungal cells may be included in the debris and not reached by lytic enzymes. See comments at page 9.</p> <p>Check the titre of the pathogen by plating and increase the titre for inoculation.</p> <p>See troubleshooting guides of procedures in laboratory manuals or these kits. Alternatively, use Molzym's complete DNA isolation kits which have been extensively evaluated for isolation of pico to femtogram amounts of pathogen DNA.</p>

	<ul style="list-style-type: none"> • Wrong elution conditions • Loss of nucleic acids during the storage of the eluate 	<p>For elution of the DNA from the column matrix, good experience was made using DNA-free water, PCR grade (Molzym order no. P-020-0003) heated to 70°C. This increases the DNA yield significantly.</p> <p>Store the eluted DNA at +4 to +12°C if analysed at the same day or freeze at -15 to -25°C until further use. Avoid frequent freeze-thaw cycles, because this may result in a decay of the eluted DNA (in particular at low DNA concentrations).</p>
<p>False positive PCR result</p>	<ul style="list-style-type: none"> • Cross contamination • Contamination during handling 	<p>Principally, work under UV-irradiated workstations. Avoid the generation of aerosols by careful pipetting. Frequently change gloves. UV-irradiate the workstation at the end of handling a series of samples. Prepare mastermixes and handle DNA samples under different UV workstations (page 7). Use DNA-free pipette tips and other plastics. Make sure that the other Mini kit used for nucleic acid extraction and purification is DNA-free.</p>

References

- Benítez-Páez A, Álvarez M, Belda-Ferre P, Rubido S, Mira A, Tomás I** (2013) Detection of transient bacteraemia following dental extractions by 16S rDNA pyrosequencing: A pilot study. *PLoS ONE* **8**, e57782. doi:10.1371/journal.pone.0057782.
- Bille J** (2010) New nonculture-based methods for the diagnosis of invasive candidiasis. *Curr Opin Crit Care* **16**, 460-464.
- Brule AJC** (2011) Acceleration of the direct identification of *Staphylococcus aureus* versus coagulase-negative staphylococci from blood culture material: a comparison of six bacterial DNA extraction methods. *Eur J Clin Microbiol Infect Dis* **30**, 337-342.
- Disqué C** (2007) Einfluss der DNA-Extraktion auf die PCR-Detektion von Sepsiserregern. *BIOSpektrum* **06**, 627-629.
- Downey LC, Smith BP, Benjamin DK, Cohen-Wolkowicz** (2010) Recent advances in the detection of neonatal candidiasis. *Curr Fungal Infect Rep* **4**, 17-22.
- Esteban J, Alonso-Rodríguez N, del-Prado G, Ortiz-Pérez A, Molina-Manso D, Cordero-Ampuero J, Sandova E, Fernández-Roblas R, Gómez-Barrena E** (2012) PCR-hybridization after sonication improves diagnosis of implant-related infection. *Acta Orthopaedica* **3**, 299-304.
- Gebert S, Siegel D, Wellinghausen N** (2008) Rapid detection of pathogens in blood culture bottles by real-time PCR in conjunction with the pre-analytic tool MoYsis. *J Infect* **57**, 307-316.
- Handschr M, Karlic H, Hertl C, Pfeilstöcker M, Haslberger AG** (2009) Preanalytic removal of human DNA eliminates false signals in general 16S rDNA PCR monitoring of bacterial pathogens in blood. *Comp Immunol Microbiol Infect Dis* **32**, 207-219.
- Hansen WLJ, Bruggeman CA, Wolffs PFG** (2009) Evaluation of new preanalysis sample treatment tools and DNA isolation protocols to improve bacterial pathogen detection in whole blood. *J Clin Microbiol* **47**, 2629-2631.
- Hansen WLJ, Bruggeman CA, Wolffs PFG** (2013) Pre-analytical sample treatment and DNA extraction protocols for the detection of bacterial pathogens from whole blood. *Meth Mol Biol* **943**, 81-90.
- Horz HP, Scheer S, Huenger F, Vianna ME, Conrads G** (2008) Selective isolation of bacterial DNA from human clinical specimens. *J Microbiol Meth* **72**, 98-102.
- Horz HP, Scheer S, Vianna ME, Conrads G** (2010) New methods for selective isolation of bacterial DNA from human clinical specimens. *Anaerobe* **16**, 47-53.
- Kemp M, Jensen KH, Dargis R, Christensen JJ** (2010) Routine ribosomal PCR and DNA sequencing for detection and identification of bacteria. *Future Microbiol* **5**, 1101-1107.
- Laakso S, Mäki M** (2013) Assessment of a semi-automated protocol for multiplex analysis of sepsis-causing bacteria with spiked whole blood samples. *MicrobiologyOpen* **2**, 284-292.
- Leggieri N** (2010) Molecular diagnosis of bloodstream infections: planning to (physically) reach the bedside. *Curr Opin Infect Dis* **23**, 311-319.
- Loonen AJM, Jansz AR, Kreeftenberg H, Bruggeman CA, Wolffs PFG, van den Mayr A, Lass-Flörl C** (2011) Non-culture-based methods for the diagnosis of invasive candidiasis. *Curr Fungal Infect Reports* **5**, 151-156.
- Loonen AJM, Jansz AR, Stalpers J, Wolffs PFG, van den Brule AJC** (2012) An evaluation of three processing methods and the effect of reduced culture times for faster direct identification of pathogens from BacT/ALERT blood cultures by MALDI-TOF MS. *Eur J Clin Microbiol Infect Dis* **31**, 1575-1583.

- McCann CD, Jordan JJ** (2014) Evaluation of MolYsis™ Complete5 DNA extraction method for detecting *Staphylococcus aureus* DNA from whole blood in a sepsis model using PCR/pyrosequencing. *J Microbiol Meth* **99**, 1-7.
- Meurs KM, Heaney AM, Atkins CE, DeFrancesco TC, Fox PR, Keene BW, Kelliham HB, Miller MW, Oyama MA, Oaks JL** (2011) Comparison of polymerase chain reaction with bacterial 16S primers to blood culture to identify bacteremia in dogs with suspected bacterial endocarditis. *J Vet Inter Med* **25**, 959–962.
- Richardson LJ, Kaestli M, Mayoa M, Bowers JR, Tuanyok A, Schupp J, Engelthaler D, Wagner DM, Keim PS, Currie BJ** (2012) Towards a rapid molecular diagnostic for melioidosis: Comparison of DNA extraction methods from clinical specimens. *J Microbiol Meth* **88**, 179–181.
- Richtlinie über die ordnungsgemäße Entsorgung** von Abfällen aus Einrichtungen des Gesundheitsdienstes vom 01.01.2002, Robert-Koch-Institut
- Rudkjøbing VB, Aanaes K, Wolff TY, von Buchwald C, Johansen HK, Thomsen TR** (2014) An exploratory study of microbial diversity in sinus infections of cystic fibrosis patients by molecular methods. *J Cys Fibrosis* doi:10.1016/j.jcf.2014.02.008.
- Thomsen TR, Xu Y, Lorenzen J, Nielsen PH, Schønheyder HC** (2012) Improved diagnosis of biofilm infections using various molecular methods. *Culture negative orthopedic biofilm infections*. Springer Series on Biofilms **7**, 29-41.
- Wellinghausen N, Siegel D, Gebert S, Winter J** (2009) Rapid detection of *Staphylococcus aureus* bacteremia and methicillin resistance by real-time PCR in whole blood samples. *Eur J Clin Microbiol Infect Dis* **28**, 1001-1005.
- Wellinghausen N, Siegel D, Winter J, Gebert S** (2009) Rapid diagnosis of candidaemia by real-time PCR detection of *Candida* DNA in blood samples. *J Med Microbiol* **58**, 1106-1111.
- Wiesinger-Mayr H, Jordana-Lluch E, Martró E, Schoenthaler S, Noehammer C** (2011) Establishment of a semi-automated pathogen DNA isolation from whole blood and comparison with commercially available kits. *J Microbiol Meth* **85**, 206–213.
- Wolff TY, Moser C, Bundgaard H, Høiby N, Nielsen PH, Thomsen TR** (2011) Detection of microbial diversity in endocarditis using cultivation-independent molecular techniques. *Scand J Infect Dis* **43**, 857-869.
- Xu Y, Børsholt Rudkjøbing V, Simonsen O, Pedersen C, Lorenzen J, Schønheyder HC, Nielsen PH, Rolighed Thomsen T** (2012) Bacterial diversity in suspected prosthetic joint infections: an exploratory study using 16S rRNA gene analysis. *FEMS Immunol Med Microbiol* **65**, 291–304.
- Zhou L, Pollard AJ** (2012) A novel method of selective removal of human DNA improves PCR sensitivity for detection of *Salmonella* Typhi in blood samples. *BMC Inf Dis* **12**, 164 doi:10.1186/1471-2334-12-164.

Information *MolYsis*[™] Complete Kits and DNA-Free PCR Reagents

MolYsis[™] Complete

The *MolYsis*[™] Complete Kits (*MolYsis*[™] Complete 5 and *MolYsis*[™] Complete 10) are the complete solution for the human DNA removal and the microbial DNA isolation.

The kits provide reagents and materials for human DNA removal, bacteria and fungi enrichment and microbial DNA isolation. *MolYsis*[™] Complete5 enables midi size preparations from ≤1ml and 5ml volumes. Maxi size isolation of enriched microbial DNA from 5-10ml volumes is possible with *MolYsis*[™] Complete10. The kits guarantee exclusion of false signals by unspecific background amplification thus constituting your standard for extremely high PCR detection sensitivity: Enhancement over total DNA extraction can account for up to 40,000-fold.

DNA-Free PCR Reagents

A common drawback of PCR assays targeting bacterial sequences is the contamination of amplification reagents by bacterial DNA. This problem becomes even more evident when the assay is directed to a broad range of bacteria. Consequences of DNA contamination may be false-positive results and loss in analytical sensitivity.

Molzym's Mastermix 16S and other PCR reagents are guaranteed free of contaminating DNA thus generating reliable results.

Products offered include DNA-free Taq polymerase (*MolTaq 16S*) and various ready-to-use mastermixes for the detection of femtogram amounts of bacterial DNA. Amplification reactions can be performed over 40 cycles. DNA-free *MolTaq 16S* is a highly active Taq DNA polymerase for ultra-sensitive PCR analysis of bacterial DNA in samples.

The mastermixes contain all reagents for optimal amplification: dNTPs, buffer, magnesium ions (3mM final) and bovine serum albumin. If you want to run your specific assays, just add primers to *Mastermix 16S Basic* and *Mastermix 16S Dye* and start the PCR and Real-Time PCR (SYBR Green) reaction, respectively. Complete assays are available with *Mastermix 16S Primer* and *Mastermix 16S Complete* (V3/V4 region of the 16S rRNA gene). Further information see Molzym's homepage (www.molzym.com).

DNA-Free PCR Product order information

Product	Contents	Cat. No.
<i>Mastermix 16S Complete</i>	100 reactions	S-020-0100
Universal 16S rDNA PCR and Real-Time PCR assay for detection of bacteria	250 reactions 1000 reactions	S-020-0250 S-020-1000
<i>Mastermix 16S Primer</i>	100 reactions	S-021-0100
PCR assay for universal PCR detection of bacteria	250 reactions 1000 reactions	S-021-0250 S-021-1000
<i>Mastermix 16S Dye</i>	100 reactions	S-030-0100
Premixed reagents and fluorescent dye for Real-Time PCR with custom primers	250 reactions 1000 reactions	S-030-0250 S-030-1000
<i>Mastermix 16S Basic</i>	100 reactions	S-040-0100
Premixed reagents for PCR analysis with custom primers	250 reactions 1000 reactions	S-040-0250 S-040-1000
<i>MolTaq 16S</i>	100 units	P-019-0100
Taq DNA polymerase, DNA-free	500 units	P-019-0500
<i>DNA-free water, PCR grade</i>	10x 1.7ml	P-020-0003

Technical Support

If you have questions please contact us.

Our hotline: +49(0)421 69 61 62 0 • **E-Mail:** support@molzym.com • **Web:** www.molzym.com

Material safety data sheets are available on request.

Order Information

Product	Contents	Cat. No.
<i>MolYsis™ Basic10</i>	50 sample pre-treatments: Human/animal DNA removal, pathogen cell wall degradation	D-305-050
	100 sample pre-treatments: Human/animal DNA removal, pathogen cell wall degradation	D-305-100

Related Products for Whole Blood and other Liquid Samples

Product	Contents	Cat. No.
Pre-treatment of samples of small and medium sizes (used with other DNA isolation kits)		
<i>MolYsis™ Basic</i>	50 reactions	D-300-050
0.2ml sample volumes	100 reactions	D-300-100
<i>MolYsis™ Basic5</i>	50 reactions	D-301-050
≤1ml and 5ml sample volumes	100 reactions	D-301-100
Complete manual system of sample pre-treatment, extraction and DNA purification		
<i>MolYsis™ Complete5</i>	50 reactions	D-321-050
≤1ml and 5ml sample volumes	100 reactions	D-321-100
<i>MolYsis™ Complete10</i>	50 reactions	D-325-050
5 to 10ml sample volumes	100 reactions	D-325-100

See also Molzym's homepage (www.molzym.com) for automated pathogen DNA isolation products and highly active, DNA-free Taq polymerase, mastermixes and 16S rRNA gene PCR assays.

Order Hotline:

Tel.: +49(0)421 69 61 62 0 • **Fax:** +49(0)421 69 61 62 11 • **E-Mail:** order@molzym.com

Contact

Molzym GmbH & Co. KG

Mary-Astell Str. 10

D-28359 Bremen, Germany

Tel.: +49(0)421 69 61 62 0 • **Fax:** +49(0)421 69 61 62 11

E-Mail: info@molzym.com • **Web:** www.molzym.com